Cleo® 90 Infusion Set Training Guide

To insert site and connect the tubing set



Wash hands, dry thoroughly.



Prepare site, allow to dry.



Twist cap to remove.



adhesive is missing

from the top of the Cleo® set. Do not use if spring has activated and is visible in the device grip area. • When Cleo® set adhesive has touched skin, do not reposition, as adhesive will lift



Inspect device for adhesive and insertion needle.



Press device firmly on site. Hold for 5 seconds, pull straight off.






Press to secure adhesive.



Connect the tubing set and fill according to the pump User Manual.



Pinch slide connector arms inward and slide fully open.



Position site connector onto site with tubing in desired



Verify needle retracted. If still visible, press down on outer edge until needle retracts fully.



inspect to see if needle is retracted; if still visible press down on outer edge until needle retracts. · Dispose of in a manner consistent with other disposable medical products.

prematurely from the device.

· Before disposal,

To disconnect

Secure connector over site,

push together until audible

10

. "click."



Pinch connector arms to open. Lift off site.

To remove site

(U100 insulin)

Fill cannula with medication.



Press finger near peel dot. Peel back adhesive.

0.2U (0.002ml

NOTE: This training guide is intended as a reference only. Please refer to the Instructions for Use provided with this product, which contains essential information, warnings and cautions of which the user must be made aware.

smiths medical

Smiths Medical MD, Inc. Phone: 1-800-426-2448 Smiths Medical, part of the global technology business Smiths Group

For more information, contact Smiths Medical MD Customer Support at 1-800-826-9703 or visit us on the web at www.cleoinfusionsets.com